

DECLARATION OF CONFORMITY

in relation to

DIRECTIVE 2002/95/EC OF THE EUROPEAN PARLIAMENT
AND OF THE COUNCIL
OF 27 JANUARY 2003
ON THE RESTRICTION OF THE
USE OF CERTAIN HAZARDOUS SUBSTANCES (RoHS) IN
ELECTRICAL AND ELECTRONIC EQUIPMENT

Magnet Power Supplies (MPS)

Lake Shore Magnet Power Supplies are used to supply current to laboratory electromagnets and other resistive loads. The superconducting magnet power supply is used to supply current to superconducting magnets. When integrated into magnetic field systems, they facilitate a wide range of applications including, but not limited to, magneto-optical studies, magnetic hysteresis and susceptibility, and Hall effect measurements, as well as in-line annealing to study how the properties of materials change at high temperatures.

The equipment detailed above is classified as category 9 equipment 'Monitoring and control instruments' as defined within Directive 2002/96/EC on Waste Electrical and Electronic Equipment (WEEE). As such, the equipment falls outside the scope of Directive 2002/96/EC on the Restriction of the use of certain Hazardous Substances (RoHS) and may legitimately contain substances restricted by Article 4.1 of Directive 2002/95/EC within each homogeneous element above the maximum concentration values defined within Commission Decision 2005/618/EC.

On behalf of Lake Shore: Karen E Lint
Karen E. Lint, COO

Date: 5/23/2006



Lake Shore Cryotronics Inc. • 575 McCorkle Boulevard • Westerville, OH 43082
Sales: (614) 891-2244 • Fax: (614) 891-1392 • sales@lakeshore.com • www.lakeshore.com